

EarWell™

Infant Ear Correction System

Parent Instructions for use and care

The contained information will assist in maximizing the successful results from your child's EarWell™ course of therapy.

Please read all information and carefully follow all instructions.

U.S. and International patents pending

Becon Medical Limited
2307, Brinmore Court,
Naperville, IL-60540
630-357-7374
630-428-3816 Fax
www.beconmedical.com

©Copyright Becon Medical Limited
48BM0008Rev.B

Your child's physician has applied the EarWell™ ear correction system. At the conclusion of the therapy your baby's ears should be properly formed.

During the course of therapy there are several things that you should know to enhance the results of the EarWell™ application.

ADHESIVE

- EarWell is affixed to your baby's skin by a special medical adhesive designed to maximize adhesion and minimize trauma. It will hold the EarWell™ reliably in place for approximately fourteen to seventeen days.
- Inspect the cradle daily to verify its continued adhesive skin bond. When you notice that the cradle adhesive begins to separate from the skin, call your physician to schedule an appointment. During the appointment a new set of EarWell™ cradles will be re-applied.
- After phoning your physician for an appointment use the provided roll of tape to temporarily secure the EarWell™ to the skin. Cut a small piece of tape from the roll (3 inches or less) and apply over the top of the EarWell™ and then onto the skin of your child. If the area of separation is larger than the width of the tape, use additional pieces of tape to secure it to the skin.
- There are also occasions when one or both EarWell™ cradles prematurely separate from the skin. This can occur when a child becomes unusually active or restless resulting in vigorous movement of arms and shoulders. Upon observing such separation use the tape to secure the cradle to the skin in the manner described above.
- Contact your physician upon securing the EarWell™ with tape.
- If the EarWell™ becomes completely dislodged, or for more than 50% of the EarWell™ surface area, DO NOT attempt to tape it back in place. Immediately call your physician to schedule an appointment.

- It is important to note that the EarWell™ becoming temporarily dislodged will not affect the results or the duration of the therapy.

- Simply schedule an appointment for as soon as possible.

WASHING

- Because adhesives are an intricate part of the EarWell™ attachment, it should remain as dry as possible.
- It is recommended that you do not wet the EarWell™ or attempt to wash it.

REDNESS

- EarWell™ is made of a clear material to permit ear visibility.
- If the ear becomes unusually red in colour, report it to the applying physician.

CRADLE MAINTENANCE

- Care should be given to keeping the cradle interior clear of milk, formula or vomit. If any of these fluids penetrate to the inside of the cradle, make an appointment with the applying physician for removal of the top shell and cleaning of the ear and the EarWell™ components.
- Particular attention should be devoted to protecting the cradle interior from the presence of any vomit.

COVER REMOVAL

- The top cover of the EarWell™ should be removed by a physician.

EarWell™

Sistema correctivo de orejas infantiles

Padres Instrucciones de Uso y Cuidado

La siguiente información ayudará a maximizar los resultados exitosos de la terapia EarWell™ para su bebé.

Por favor, lea toda la información y siga todas las instrucciones con cuidado.

Patentes estadounidense e internacional pendientes

Becon Medical Limited
2307, Brinmore Court,
Naperville, IL-60540
630-357-7374
630-428-3816 Fax
www.beconmedical.com

©Copyright Becon Medical Limited
48BM0008Rev.B

El médico de su hijo o hija ha aplicado el sistema correctivo EarWell™. Al final de la terapia las orejas de bebé deben estar bien formadas.

Durante el curso de la terapia hay unas cosas que se deben saber para mejorar los resultados de la aplicación del EarWell™.

ADHESIVO

- El EarWell™ está pegado a la piel de su bebé por un adhesivo médico diseñado para maximizar la adhesión y minimizar el trauma. Mantendrá el EarWell en su lugar para aproximadamente catorce a diecisiete días.
- Inspeccione la cuna diariamente para verificar que el adhesivo esté pegado a la piel. Cuando vea que el adhesivo deja de pegarse a la piel, llame al médico para hacer una cita. Durante esa cita unas cunas nuevas de EarWell™ serán aplicadas.
- Después de llamar a su médico, use el rollo de cinta Micropore para fijar el EarWell™ a la piel. Corte un trozo pequeño de cinta del rollo (3 pulgadas o menos) y aplique encima del EarWell y de la piel del bebé. Si el área de separación está más grande que el ancho de la cinta, use más cinta para asegurar el EarWell™ a la piel.
- También hay ocasiones cuando una o ambas cunas EarWell™ se separan de la piel prematuramente. Esto puede ocurrir cuando un niño se pone excepcionalmente activo o inquieto, resultando en movimiento enérgico de los hombros y brazos. Al observar tal separación, use la cinta Micropore para asegurar la cuna a la oreja como indicado arriba.
- Llame al médico cuando asegure el EarWell™ con la cinta.
- Si el EarWell™ se separa completamente, o más de 50% del área de la superficie de aparato, NO INTENTES asegurarlo de nuevo con cinta. Llame a su médico para hacer una cita.

- Es importante notar que si el EarWell™ se separa temporariamente no afectará los resultados ni la duración de la terapia.

- Simplemente haga una cita para lo más pronto posible.

LAVAR

- Porque el adhesivo es una parte intrincada del EarWell™, debe mantenerlo el más seco posible.
- Se recomienda que no intente mojar ni lavar el EarWell.

ENROJECIMIENTO

- EarWell™ está hecho de un material transparente para que la piel sea visible.
- Si la oreja se pone más roja de lo normal, avise al médico.

MANTENIMIENTO DE LA CUNA

- El interior de la cuna debe mantenerse libre de leche, fórmula o vómito. Si alguno de estos fluidos penetra al interior de la cuna, solicite una cita con el médico tratante para retirar la cubierta superior y limpiar el oído y los componentes EarWell™.
- Se debe prestar especial atención para proteger el interior de la cuna de la presencia de vómito.

TAPA

- Solo el médico debe quitar la tapa del EarWell™.